

**PLASTIC WELDING**

Leister Technologies AG

PROCESS HEAT

Galileo-Strasse 10
CH-6056 Kaegiswil

LASERSYSTEMS

phone +41 41 662 74 74
fax +41 41 662 74 16

leister@leister.com
www.leister.com

Leister Air Pressure Testing Device (150.720)

Wika Manometer AG confirms that the standard air pressure gauge 40mm, metal housing double scale 0-6 Bar / 0-87 PSI class 1.6 conforms to EN 837 and that the Manometer is manufactured to hold an accuracy tolerance of +/- 1.6%.

Signed,

Leister Technologies AG

Patrick Schnider
Product Manager

Date: 01.12.2020